

# Programat® P710

## Software Revision Log

Version	Changes from previous version	Status
<b>V3.1</b> 01.02.2018	<b>Software optimized</b> <ul style="list-style-type: none"> <li>WLAN performance has been enhanced</li> <li>Push function can be tested</li> <li>Multimedia files can be directly deleted</li> <li>Various languages have been improved</li> <li>Various software optimizations</li> </ul>	<b>Currently delivered version</b>
<b>V3.0</b> 28.04.2017	<b>Software optimized</b> <ul style="list-style-type: none"> <li>Now with IRT2 function that shows the ideal time when to remove the fired object from the furnace</li> <li>New program structure with new program parameters for IPS Style, IPS e.max Press Layering Technique, IPS e.max ZirPress Layering Technique, IPS e.max ZirCAD Staining Technique, IPS e.max CAD Crystall/Glaze, IPS e.max CAD Layering Technique, IPS e.max CAD Layering Technique, IPS d.SIGN, IPS e.max ZirCAD Layering Technique</li> <li>Optimized Vacuum test program</li> <li>Various software optimizations</li> </ul>	<b>No longer available!</b>
<b>V2.0</b> 08.04.2016	<b>Software optimized</b> <ul style="list-style-type: none"> <li>New programs for IPS Style and IPS Ivocolor</li> <li>DSA - Digital Shade Assistant: Orientation of the reference teeth either "incisal edge to incisal edge" or "incisal edge to tooth neck" possible</li> <li>In-process dehumidification: Function that protects the vacuum system from contamination and its consequential damages</li> <li>Various software optimizations</li> </ul>	<b>No longer available!</b>
<b>V1.0</b> 15.06.2015	<b>Basic version</b>	<b>No longer available!</b>

Version 01.02.2018